

for research use only

Certificate of Analysis

Catalog No: AAA10550	
Product Name: Qualitative Human Yellow Fever Virus Antibody IgG (YFV-IgG) ELISA Kit	
Date of Production: December 28, 2023	Shelf-Life of Product: Six Months (2°C-8°C)
Specification: 96T	Quantity: 2
Testing Date: December 30, 2023	Report Date: December 30, 2023

Inspection item	Description				Result
Components	Items	Materials	48 Tests	96 Tests	Intact Correct Complete
	1	Plate	48 well plate	96 well plate	
	2	Controls (2 vial)	0.5ml×2 vials	0.5ml×2 vials	
	3	Specimen Diluent	3.0ml	6.0ml	
	4	HRP-Conjugate Reagent	5.0ml	10.0ml	
	5	20×Wash Solution	15ml	25ml	
	6	Stop Solution	3.0ml	6.0ml	
	7	Chromogen Solution A	3.0ml	6.0ml	
	8	Chromogen Solution B	3.0ml	6.0ml	
	9	Closure Plate Membrane	2	2	
	10	Sealed Bags	1	1	
	11	Instruction	1	1	
Label	All Labels				Intact, Correct, Complete
Package	All Packages				Intact, Correct, Complete
Inter-assay CV%:	≤15%				Yes
Intra-assay CV%:	≤15%				Yes
Conclusion	Qualified products				