

for research use only

## CERTIFICATE OF ANALYSIS

Human Anti-diuretic hormone/vasopressin/arginine vasopressin,

ADH/VP/AVP ELISA Kit

Catalog: AAA11280

### TEST PARAMETER

MATERIALS SUPPLIED	Standard solution	160ng/L	approved
	Biotinylated human ADH/VP/AVP Antibody	1ml	approved
	Washing Buffer(25X)	20ml	approved
	Streptavidin-HRP	6ml	approved
	Pre-coated ELISA Plate	12-well * 8 tubes	approved
	Standard Diluent	3ml	approved
	Substrate solution A	6ml	approved
	Substrate solution B	6ml	approved
	Stop solution	6ml	approved
CV(%)=SD/meanX100	Intra-Assay: CV<8%	Inter-Assay: CV<10%	approved
ASSAY RANGE	0.5ng/L - 150ng/L		approved
SENSITIVITY	0.28ng/L		approved
LOT #	201911010		approved
QC	Conclusion: Approved		
	Signature/Date:2019.10.27		