for research use only

CERTIFICATE OF ANALYSIS

Product Name	Human cluster Of differentiation 4 ELISA KIT		
Item No.	AAA11292		
TEST PARAMETER			
	Standard solution	32ng/ml	approved
	Biotinylated Antibody	1ml	approved
	Washing concentrate (25X)	20ml	approved
MATERIALS SUPPLIED	Streptavidin-HRP	6ml	approved
MFG.:2024/08/26 EXP.:2025/02/25 Lot:202408018	Pre-coated ELISA plate	12 * 8 well strips x1	approved
	Standard diluent	3ml	approved
	Substrate Solution A	6ml	approved
	Substrate Solution B	6ml	approved
	Stop solution	6ml	approved
CV(%)=SD/meanX100	Intra-Assay: CV<10%	Inter-Assay: CV<12%	approved
ASSAY RANGE	0.1-30ng/ml		approved
SENSITIVITY	0.05ng/ml		approved
	Conclusion: approved		
	Date:2024/08/26		