

CERTIFICATE OF ANALYSIS

Product Name	Human Endogenous secretory receptor for Advanced Glycation end Products ELISA KIT		
Item No.	AAA11369		
TEST PARAMETER			
MATERIALS SUPPLIED MFG.:2023/09/11 EXP.:2024/03/11 Lot:202309007	Standard solution	1600ng/L	approved
	Biotinylated Antibody	1ml	approved
	Washing concentrate (25X)	20ml	approved
	Streptavidin-HRP	6ml	approved
	Pre-coated ELISA plate	12 * 8 well strips x1	approved
	Standard diluent	3ml	approved
	Substrate Solution A	6ml	approved
	Substrate Solution B	6ml	approved
	Stop solution	6ml	approved
CV(%)=SD/meanX100	Intra-Assay: CV<10%	Inter-Assay: CV<12%	approved
ASSAY RANGE	5-1500ng/L		approved
SENSITIVITY	2.48ng/L		approved
	Conclusion: approved		
	Date:2023/09/11		