

## Certificate of Analysis

COA No:20220127514

Catalog	<b>AAA13052</b>
Date	2022/01/27
Lot	36396217
Quantity	1Kit

This is to certify that we, the undersigned, have inspected the quality of the above mentioned kit and found the result of inspection as follows.

Items	Result
Toxicity	None
Specific	No cross-reaction with other factors
Relative standard curve coefficient	0.9983
Intra assay Precision	$\leq 8\%$
Recovery	81%