

Certification Of Analysis

Product Name: Pig Hyaluronic acid (HA) ELISA Kit

Catalog Number:'AAA13239

Species Reactivity: Pig (*Sus scrofa*; Porcine)

Lot No: L05WHA29 **Mfg:** May 29, 2023 **Exp:** Nov 28, 2023

Introduction

Item	Standard		Test Result
Description	For the quantitative detection of Pig Hyaluronic acid (HA) concentration in serum, plasma and other biological fluids.		Conform
Identification	Sandwich		Positive
Composition	Assay plate (96 Wells)	1	Conform
	Standard (lyophilized)	2	
	Sample Diluent	1 × 20 mL	
	Biotin-Conjugate (concentrate 100 x)	1 × 120 µL	
	Biotin-Conjugate Diluent	1 × 20 mL	
	Streptavidin-HRP (concentrate 100 x)	1 × 120 µL	
	Streptavidin-HRP Diluent	1 × 20 mL	
	Substrate Solution	1 × 12 mL	
	Stop Solution	1 × 10 mL	
	Wash Buffer (concentrate 25 x)	1 × 20 mL	
	Adhesive Films	4	Conform
	Instruction manual	1	
Assay Range	7.8-500 ng/mL		Conform


Sensitivity

The limit of detection of Pig HA defined as the analyte concentration resulting in an absorbance significantly higher than that of the dilution medium (mean plus 2 standard deviations) was determined to be 3.9 ng/mL (mean of 6 independent assays).

Typical data

For convenience in result calculation, absorbance as abscissa and standard concentrations can be used as ordinate. The standard curve provided in the manual is only for reference, experimenters should draw the standard curve based on data of themselves.

ng/mL	Standard		Average
500	2.623	2.605	2.614
250	1.994	2.012	2.003
125	1.493	1.481	1.487
62.5	0.990	1.002	0.996
31.25	0.622	0.628	0.625
15.6	0.386	0.372	0.379
7.8	0.252	0.242	0.247
0	0.173	0.199	0.186



4th Degree Polynomial Fit: $y=a+bx+cx^2+dx^3+ex^4$

Coefficient Data:

a = -1.66266077766E+001

b = 1.08539160444E+002

c = -8.06586625621E+001

d = 5.44453702622E+001

e = -4.03502210428E+000

Recovery

The recovery of Pig HA spiked to levels throughout the range of the assay was evaluated.

Sample Type	Number	Recovery range (%)	Average(%)
Human serum	10	91-99	95
Human EDTA plasma	10	89-101	95

Linearity

To assess the linearity of the assay, samples containing high concentrations of Pig HA were serially diluted with Calibrator Diluent to produce samples with values within the dynamic range of the assay.

Sample Type	1: 2	1: 4	1: 8	1: 16
Human serum	91-102%	89-100%	91-97%	92-101%
Human EDTA plasma	88-96%	92-99%	91-101%	90-103%

Precision

Intra-assay Precision (Precision within an assay)

Three samples of known concentration were tested twenty times on one plate to assess intra-assay precision.

Inter-assay Precision (Precision between assays)

Three samples of known concentration were tested in forty separate assays to assess inter-assay precision.

$CV (\%) = SD/mean \times 100$

	Intra-assay Precision			Inter-assay Precision		
Sample	1	2	3	1	2	3
n	20	20	20	20	20	20
Mean (ng/mL)	0.398	0.799	1.016	0.425	0.864	1.148
SD	0.027	0.058	0.078	0.031	0.069	0.101
CV (%)	6.8	7.3	7.7	7.3	8.0	8.8

Approved Date: 2023.05.29