

## Certification Of Analysis

**Product Name:** Human Bone alkaline phosphatase (BALP) ELISA Kit

**Catalog Number:** AAA13282

**Species Reactivity:** Human (Homo sapiens)

**Lot No:** L10WBA11    **Mfg:** Oct 11, 2023    **Exp:** Apr 10, 2024

### Introduction

Item	Standard		Test Result
Description	For the quantitative detection of Human Bone alkaline phosphatase (BALP) concentration in serum, plasma and other biological fluids.		Conform
Identification	Sandwich		Positive
Composition	Assay plate (96 Wells)	1	Conform
	Standard (lyophilized)	2	
	Sample Diluent	1 × 20 mL	
	Biotin-Conjugate (concentrate 100 x)	1 × 120 µL	
	Biotin-Conjugate Diluent	1 × 20 mL	
	Streptavidin-HRP (concentrate 100 x)	1 × 120 µL	
	Streptavidin-HRP Diluent	1 × 20 mL	
	Substrate Solution	1 × 12 mL	
	Stop Solution	1 × 10 mL	
	Wash Buffer (concentrate 25 x)	1 × 20 mL	
	Adhesive Films	4	Conform
	Instruction manual	1	
Assay Range	0.156 - 10 ng/mL		Conform


### Sensitivity

The limit of detection of Human BALP defined as the analyte concentration resulting in an absorbance significantly higher than that of the dilution medium (mean plus 2 standard deviations) was determined to be 0.1 ng/mL (mean of 6 independent assays).

### Typical data

For convenience in result calculation, absorbance as abscissa and standard concentrations can be used as ordinate. The standard curve provided in the manual is only for reference, experimenters should draw the standard curve based on data of themselves.

ng/mL	Standard		Average
200	2.618	2.630	2.624
100	1.983	2.001	1.992
50	1.451	1.427	1.439
25	0.886	0.928	0.907
12.5	0.605	0.629	0.617
6.25	0.427	0.435	0.431
3.12	0.241	0.259	0.250
0	0.111	0.087	0.099



3rd degree Polynomial Fit:  $y=a+bx+cx^2+dx^3...$

Coefficient Data:

a = -4.87402554483E+000

b = 3.77957069312E+001

c = -1.72913623912E+001

d = 1.24347612754E+001

## Recovery

The recovery of Human BALP spiked to levels throughout the range of the assay was evaluated.

Sample Type	Number	Recovery range (%)	Average(%)
Human serum	10	91-98	95
Human plasma	10	89-97	93

## Linearity

# for research use only

To assess the linearity of the assay, samples containing high concentrations of Human BALP were serially diluted with Sample Diluent to produce samples with values within the dynamic range of the assay.

Sample Type	1: 2	1: 4	1: 8	1: 16
Human serum	90-96%	89-94%	92-99%	94-101%
Human plasma	89-96%	90-95%	92-98%	92-99%

## Precision

Intra-assay Precision (Precision within an assay)

Three samples of known concentration were tested twenty times on one plate to assess intra-assay precision.

Inter-assay Precision (Precision between assays)

Three samples of known concentration were tested in forty separate assays to assess inter-assay precision.

$CV (\%) = SD/mean \times 100$

	Intra-assay Precision			Inter-assay Precision		
Sample	1	2	3	1	2	3
n	20	20	20	20	20	20
Mean (ng/mL)	0.344	0.605	0.927	0.275	0.663	1.194
SD	0.023	0.044	0.070	0.020	0.050	0.101
CV (%)	6.7	7.3	7.6	7.3	7.5	8.5

**Date:** 2023.10.11