

AAA15341

**PRECISION**

**Intra-assay Precision (Precision within an assay): CV%<8%**


Three samples of known concentration were tested twenty times on one plate to assess.

**Inter-assay Precision (Precision between assays): CV%<10%**

Three samples of known concentration were tested in twenty assays to assess.

**TYPICAL DATA**

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



| pg/ml | OD1   | OD2   | Average | Corrected |
|-------|-------|-------|---------|-----------|
| 2500  | 2.441 | 2.382 | 2.412   | 2.301     |
| 1250  | 1.692 | 1.725 | 1.709   | 1.598     |
| 625   | 1.021 | 1.103 | 1.062   | 0.951     |
| 312.5 | 0.563 | 0.539 | 0.551   | 0.440     |
| 156   | 0.327 | 0.314 | 0.321   | 0.210     |
| 78    | 0.245 | 0.258 | 0.252   | 0.141     |
| 39    | 0.185 | 0.191 | 0.188   | 0.077     |
| 0     | 0.108 | 0.113 | 0.111   |           |

**LOD**

9.75pg/ml

# for research use only

## **LINEARITY**

To assess the linearity of the assay, samples were spiked with high concentrations of human GR- $\beta$  in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.

|     | Sample    | Serum(n=4) |
|-----|-----------|------------|
| 1:1 | Average % | 103        |
|     | Range %   | 99-108     |
| 1:2 | Average % | 98         |
|     | Range %   | 94-102     |
| 1:4 | Average % | 91         |
|     | Range %   | 88-95      |
| 1:8 | Average % | 89         |
|     | Range %   | 85-96      |

## **RECOVERY**

The recovery of human GR- $\beta$  spiked to levels throughout the range of the assay in various matrices was evaluated. Samples were diluted prior to assay as directed in the Sample Preparation section.

| Sample Type       | Average % Recovery | Range  |
|-------------------|--------------------|--------|
| Serum (n=5)       | 95                 | 90-100 |
| EDTA plasma (n=4) | 87                 | 83-95  |