

Human Glutamate [NMDA] Receptor Subunit Epsilon-2 (NR-2) Antibody ELISA Kit

PRECISION

Intra-assay Precision (Precision within an assay): CV%<15%

Three samples of known concentration were tested twenty times on one plate to assess.

Inter-assay Precision (Precision between assays):CV%<15%

Three samples of known concentration were tested in twenty assays to assess.

TYPICAL DATA

| | | |
|------------------|---------------|-------------|
| Test parameter | specification | test result |
| Positive control | >1.0 | 1.066 |
| Negative control | <0.2 | 0.095 |
| Positive rate | 10, Positive | 100% |
| Negative rate | 10, Negative | 100% |