

Certificate of Analysis

Product name: Human Carboxymethyl lysine ELISA Kit

Catalog No.: AAA16066

Lot No: 20220926C

Exp: Mar 26,2023

Manufacture date : Sep 26,2022

Valid Period : 6 months

Storage: 2-8° C

PRECISION

Inter-assay CV%: 6.0 7.5

Intra-assay CV%: 5.1 6.9

RECOVERY

Sample type

average% recovery

range

Serum

95


87-103%

LINEARITY

		Serum
1: 2	Average % of Expected	96
	Range (%)	94-103
1: 4	Average % of Expected	90
	Range (%)	89-106
1: 8	Average % of Expected	88
	Range (%)	85-109

STANDARD CURVE for demonstration only:

	Con ng/mL	OD
A	0	1.867
B	250	1.386
C	500	0.842
D	1000	0.57
E	2500	0.299
F	5000	0.104



Date: 2022.09.26