

## Certificate of Analysis

**Product name:** Rat Bone Alkaline Phosphatase ELISA Kit

**Catalog No.:** AAA16071 **Lot No:** 20170815

**Exp:** 20180315

**Expiration data:** 6 months

**Storage:** 2-8° C

### PRECISION

Inter-assay CV%: 5.0 6.5

Intra-assay CV%: 6.9 8.7

### RECOVERY

Sample type

average% recovery

range

Serum

97


86-108%

### LINEARITY

		Serum
1: 2	Average % of Expected	95
	Range (%)	91-103
1: 4	Average % of Expected	90
	Range (%)	90-105
1: 8	Average % of Expected	92
	Range (%)	84-107
	Range (%)	

### STANDARD CURVE for demonstration only:

	Con ng/mL	OD
A	0	1.824
B	1	1.435
C	2.5	0.986
D	5	0.673
E	10	0.327
F	25	0.129



Date: 2017.08.15