

for research use only

Certificate of Analysis

Product name: Mouse Vitamin A ELISA Kit

Catalog No.: AAA16093

Lot No: 20190823C

Exp: 2020/02/23

Expiration date: 6 months

Storage: 2-8° C

PRECISION

Inter-assay CV%: 5.9 7.4

Intra-assay CV%: 4.8 5.9

RECOVERY

Sample type

average% recovery

range

Serum

97


85-109%

LINEARITY

		Serum
1: 2	Average % of Expected	97
	Range (%)	92-102
1: 4	Average % of Expected	93
	Range (%)	89-109
1: 8	Average % of Expected	86
	Range (%)	86-107
	Range (%)	

STANDARD CURVE for demonstration only:

	Con (umol/L)	OD
A	0	2.215
B	0.5	1.734
C	1	1.031
D	2.5	0.428
E	5	0.309
F	10	0.185



Date: 2019.08.23