

Certificate of Analysis

Product name: Rabbit Platelet Factor 4 ELISA Kit

Catalog No.: AAA16130

Lot No: 20230128C

Exp: Jul 28,2023

Manufacture date :Jan 28,2023

Valid Period : 6 months

Storage: 2-8° C

PRECISION

Inter-assay CV%: 6.1 7.8

Intra-assay CV%: 5.1 6.5

RECOVERY

Sample type

average% recovery

range

Serum

97


89-105%

LINEARITY

		Serum
1: 2	Average % of Expected	97
	Range (%)	93-103
1: 4	Average % of Expected	93
	Range (%)	90-109
1: 8	Average % of Expected	90
	Range (%)	87-110

STANDARD CURVE for demonstration only:

	Con ng/ml	OD
A	0	1.883
B	1	1.176
C	2.5	0.917
D	5	0.609
E	10	0.351
F	25	0.193



Date: 2023.01.28