

Certificate of Analysis

Product name: Guinea pig Interleukin 8 ELISA Kit

Catalog No.: AAA16213 **Lot No:** 20170908

Exp: 20180308

Expiration data: 6 months

Storage: 2-8° C

PRECISION

Inter-assay CV%: 5.7 6.9

Intra-assay CV%: 6.1 7.8

RECOVERY

Sample type

average% recovery

range

Serum

95


83-107%

LINEARITY

		Serum
1: 2	Average % of Expected	94
	Range (%)	93-103
1: 4	Average % of Expected	90
	Range (%)	91-107
1: 8	Average % of Expected	87
	Range (%)	84-105
	Range (%)	

STANDARD CURVE for demonstration only:

		Con pg/ml	OD
0	A	0	2.619
1	B	50	2.182
2	C	100	1.765
3	D	250	0.941
4	E	500	0.488
5	F	1000	0.179



Date: 2017.09.08