

## Certificate of Analysis

**Product name:** Mouse Collagen Type I ELISA Kit

**Catalog No.:** AAA16214

**Lot No:** 20180611

**Exp:** 2018/12/11

**Expiration date:** 6 months

**Storage:** 2-8° C

### PRECISION

Inter-assay CV%: 6.4 7.6

Intra-assay CV%: 5.3 5.9

### RECOVERY

Sample type

average% recovery

range

Serum

96


83-104%

### LINEARITY

		Serum
1: 2	Average % of Expected	96
	Range (%)	93-103
1: 4	Average % of Expected	93
	Range (%)	87-106
1: 8	Average % of Expected	90
	Range (%)	84-107

**STANDARD CURVE** for demonstration only:

	Con ng/mL	OD
A	0	2.108
B	25	1.726
C	50	1.391
D	100	1.034
E	250	0.509
F	500	0.247



Date: 2018.06.11