

Certificate of Analysis

Product name: Rabbit Anti-Mullerian Hormone (AMH) ELISA Kit

Catalog No.: AAA16223 **Lot No:** BG190125RBC **Exp:** 2019/07/25

Expiration date: 6 months **Storage:** 2-8° C

PRECISION

Inter-assay CV%: 6.4 7.6 Intra-assay CV%: 5.6 7.2

RECOVERY


Sample type	average% recovery	range
Serum	97	91-103%

LINEARITY

		Serum
1: 2	Average % of Expected	97
	Range (%)	93-102
1: 4	Average % of Expected	94
	Range (%)	88-105
1: 8	Average % of Expected	88
	Range (%)	84-106

STANDARD CURVE for demonstration only:

	Con ng/ml	OD
A	0	1.979
B	1	1.752
C	2.5	1.255
D	5	0.772
E	10	0.331
F	25	0.135



Date: 2019.01.25