

Certificate of Analysis

Product name: Human N acetyl beta D glucosaminidase ELISA Kit

Catalog No.: AAA16242 **Lot No:** 20240624S **Exp:** Dec 24,2024

Manufacture date : June 24,2024 **Valid Period :** 6 months **Storage:** 2-8° C

PRECISION

Inter-assay CV%: 6.4 7.2 Intra-assay CV%: 5.6 6.3

RECOVERY


Sample type	average% recovery	range
Serum	97	91-103%

LINEARITY

		Serum
1: 2	Average % of Expected	97
	Range (%)	94-101
1: 4	Average % of Expected	93
	Range (%)	91-105
1: 8	Average % of Expected	91
	Range (%)	87-107

STANDARD CURVE for demonstration only:

	Con ng/ml	OD
A	0	0.144
B	10	0.308
C	25	0.629
D	50	1.271
E	100	1.695
F	250	2.084



Date: 2024.06.24