

## Certificate of Analysis

**Product name:** Rabbit Bradykinin ELISA Kit

**Catalog No.:** AAA16258

**Lot No:** 20191118C

**Exp:** 2020/05/18

**Expiration date:** 6 months

**Storage:** 2-8° C

### PRECISION

Inter-assay CV%: 6.9 7.1

Intra-assay CV%: 5.1 6.9

### RECOVERY

Sample type

average% recovery

range

Serum

97


85-109%

### LINEARITY

|      |                       | Serum  |
|------|-----------------------|--------|
| 1: 2 | Average % of Expected | 97     |
|      | Range (%)             | 94-102 |
| 1: 4 | Average % of Expected | 92     |
|      | Range (%)             | 89-105 |
| 1: 8 | Average % of Expected | 89     |
|      | Range (%)             | 85-109 |
|      | Range (%)             |        |

**STANDARD CURVE** for demonstration only:

|   | Con ng/ml | OD    |
|---|-----------|-------|
| A | 0         | 1.972 |
| B | 0.5       | 1.634 |
| C | 1         | 1.028 |
| D | 2.5       | 0.741 |
| E | 5         | 0.396 |
| F | 10        | 0.142 |



Date: 2019.11.18