

## Certificate of Analysis

**Product name:** Canine Coagulation Factor VIII ELISA

**Kit Catalog No.:** AAA16273    **Lot No:** 20161219    **Exp:** 20170619

**Expiration data:** 6 months    **Storage:** 2-8° C

### PRECISION

Intra-assay CV%: 5.1    6.2    Inter-assay CV%: 6.3    7.5

### RECOVERY


Sample type	average% recovery	range
Serum, Plasma	98	89-107%

### LINEARITY

		Serum
1: 2	Average % of Expected	95
	Range (%)	93-102
1: 4	Average % of Expected	93
	Range (%)	92-105
1: 8	Average % of Expected	90
	Range (%)	87-109

**STANDARD CURVE** for demonstration only:

	Con ng/ml	OD
A	0	2.013
B	5	1.678
C	10	1.209
D	25	0.732
E	50	0.391
F	100	0.186



Date: 2016.12.19