

for research use only

## Certificate of Analysis

**Product name:** Human Anti double stranded DNA ELISAKit

**Catalog No.:** AAA16285 **Lot No:** BG231110HUC

**Exp:** May 10,2024

**Manufacture date:** Nov 10,2023

**Valid Period :** 6 months

**Storage:** 2-8° C

### PRECISION

Inter-assay CV%: 6.7 7.9      Intra-assay CV%: 6.0 7.1

### RECOVERY

Sample type

average% recovery

range

Serum

96


90-102%

### LINEARITY

		Serum
1: 2	Average % of Expected	96
	Range (%)	93-102
1: 4	Average % of Expected	91
	Range (%)	90-105
1: 8	Average % of Expected	89
	Range (%)	87-110

**STANDARD CURVE** for demonstration only:

	Con ng/mL	OD
A	0	2.017
B	50	1.536
C	100	0.998
D	250	0.63
E	500	0.381
F	1000	0.173



Date: 2023.11.10