

Certificate of Analysis

Product name: Rat E2 estradiol ELISA Kit

Catalog No.: AAA16374

Lot No: 20210326C

Exp: Sep 26, 2021

Expiration date: 6 months

Manufacturing date: Mar 26, 2021

Storage: 2-8° C

PRECISION

Inter-assay CV%: 6.5 7.8

Intra-assay CV%: 4.9 5.6

RECOVERY

Sample type

average% recovery

range

Serum

97


90-107%

LINEARITY

		Serum
1: 2	Average % of Expected	97
	Range (%)	93-101
1: 4	Average % of Expected	93
	Range (%)	87-106
1: 8	Average % of Expected	89
	Range (%)	83-112

STANDARD CURVE for demonstration only:

	Con pg/mL	OD
A	0	1.859
B	250	1.275
C	500	0.942
D	1000	0.653
E	2500	0.422
F	5000	0.228



Date: 2021.03.26