

Certificate of Analysis

Product name: Human Ischemia Modified Albumin ELISA

Kit Catalog No.: AAA16444 **Lot No:** BG200323HUC **Exp:** Sep 23,2020

Expiration date: 6 months **Storage:** 2-8° C

PRECISION

Inter-assay CV%: 5.9 7.4 **Intra-assay CV%:** 4.8 5.9

RECOVERY


Sample type	average% recovery	range
Serum	99	98-100%

LINEARITY

		Serum
1: 2	Average % of Expected	97
	Range (%)	94-102
1: 4	Average % of Expected	94
	Range (%)	88-108
1: 8	Average % of Expected	89
	Range (%)	83-110
	Range (%)	

STANDARD CURVE for demonstration only:

	Con mg/ml	OD
A	0	1.924
B	5	1.563
C	10	1.178
D	25	0.607
E	50	0.294
F	100	0.113



Date: 2020.03.23