

Certificate of Analysis

Product name: Human Collagen Type III ELISA Kit

Catalog No.: AAA16536

Lot No: 20230721C

Exp: Jan 21, 2024

Manufacture date : Jul 21, 2023

Valid Period : 6 months

Storage: 2-8° C

PRECISION

Inter-assay CV%: 6.0 7.8

Intra-assay CV%: 5.2 6.3

RECOVERY

Sample type

average% recovery

range

Serum

95


92-100%

LINEARITY

		Serum
1: 2	Average % of Expected	96
	Range (%)	92-101
1: 4	Average % of Expected	92
	Range (%)	90-106
1: 8	Average % of Expected	90
	Range (%)	89-108

STANDARD CURVE for demonstration only:

	Con ng/mL	OD
A	0	1.934
B	50	1.367
C	100	0.841
D	250	0.538
E	500	0.309
F	1000	0.154



Date: 2023.07.21