

## Certificate of Analysis

**Product name:** Human Augmenter of Liver Regeneration ELISA Kit

**Catalog No.:** AAA16602

**Lot No:** 20210507S

**Exp:** Nov 07,2021

**Expiration date:** 6 months

**Manufacturing date:** May 07,2021

**Storage:** 2-8° C

### PRECISION

Inter-assay CV%: 6.9 7.8

Intra-assay CV%: 4.9 5.6

### RECOVERY

Sample type

average% recovery

range

Serum

97


86-108%

### LINEARITY

|      |                       | Serum  |
|------|-----------------------|--------|
| 1: 2 | Average % of Expected | 95     |
|      | Range (%)             | 91-102 |
| 1: 4 | Average % of Expected | 92     |
|      | Range (%)             | 89-106 |
| 1: 8 | Average % of Expected | 88     |
|      | Range (%)             | 83-108 |

**STANDARD CURVE** for demonstration only:

|   | Con ng/ml | OD    |
|---|-----------|-------|
| A | 0         | 0.141 |
| B | 0.5       | 0.366 |
| C | 1         | 0.705 |
| D | 2.5       | 1.128 |
| E | 5         | 1.636 |
| F | 10        | 1.997 |



Date: 2021.05.07