

## Certificate of Analysis

**Product name:** Sheep Interleukin 10 ELISA Kit

**Catalog No.:** AAA16701

**Lot No:** 20200430S

**Exp:** Oct 30, 2020

**Expiration date:** 6 months

**Storage:** 2-8° C

### PRECISION

Inter-assay CV%: 6.2 7.9

Intra-assay CV%: 4.9 5.3

### RECOVERY

Sample type

average% recovery

range

Serum

97


90-104%

### LINEARITY

		Serum
1: 2	Average % of Expected	97
	Range (%)	92-102
1: 4	Average % of Expected	93
	Range (%)	89-109
1: 8	Average % of Expected	88
	Range (%)	84-112
	Range (%)	

**STANDARD CURVE** for demonstration only:

	Con pg/ml	OD
A	0	0.217
B	50	0.408
C	100	0.673
D	250	1.342
E	500	1.783
F	1000	2.116



Date: 2020.04.30