

Certificate of Analysis

Product name: Bovine Heat Shock Protein 70 ELISA Kit

Catalog No.: AAA16802

Lot No: 20230707C

Exp: Jan 07,2024

Manufacture date : Jul 07,2023

Valid Period : 6 months

Storage: 2-8° C

PRECISION

Inter-assay CV%: 6.1 7.8

Intra-assay CV%: 5.2 6.6

RECOVERY

Sample type

average% recovery

range

Serum

97


92-102%

LINEARITY

		Serum
1: 2	Average % of Expected	98
	Range (%)	94-102
1: 4	Average % of Expected	93
	Range (%)	91-103
1: 8	Average % of Expected	90
	Range (%)	87-105

STANDARD CURVE for demonstration only:

	Con ng/mL	OD
A	0	1.865
B	2.5	1.334
C	5	1.057
D	10	0.768
E	25	0.336
F	50	0.122



Date: 2023.07.14