

Certificate of Analysis

Product name: Monkey Prothrombin Fragment 1+2 ELISA Kit

Catalog No.: AAA16846 **Lot No.:** 20180309

Exp: 20180909

Expiration data: 6 months

Storage: 2-8° C

PRECISION

Inter-assay CV%: 4.9 5.7 Intra-assay CV%: 5.8 6.7

RECOVERY


Sample type	average% recovery	range
Serum	97	90-104%

LINEARITY

		Serum
1: 2	Average % of Expected	95
	Range (%)	92-105
1: 4	Average % of Expected	91
	Range (%)	90-104
1: 8	Average % of Expected	87
	Range (%)	83-107

STANDARD CURVE for demonstration only:

	Con ng/mL	OD
A	0	0.107
B	2.5	0.496
C	5	0.721
D	10	1.089
E	25	1.526
F	50	1.972



Date: 2017.07.28