

Certificate of Analysis

Product name: Bovine Platelet Membrane Glycoprotein 2b3a/CD41+CD61 ELISAKit

Catalog No.: AAA16900 **Lot No:** 20161121 **Exp:** 20170521

Expiration data: 6 months **Storage:** 2-8° C

PRECISION

Intra-assay CV%: 5.3 6.5 Inter-assay CV%: 6.5 8.4

RECOVERY

Sample type	average% recovery	range
Serum, Plasma	98	90-106%

LINEARITY

		Serum
1: 2	Average % of Expected	97
	Range (%)	96-103
1: 4	Average % of Expected	93
	Range (%)	93-107
1: 8	Average % of Expected	91
	Range (%)	90-104

STANDARD CURVE for demonstration only:

	Con ng/ml	OD
A	0	1.763
B	1	1.409
C	2.5	0.984
D	5	0.719
E	10	0.451
F	25	0.264

