

Certificate of Analysis

Product name: Canine Complement 4 ELISA Kit

Catalog No.: AAA16903

Lot No: BG230217CAC

Exp: Aug 17, 2023

Manufacture date : Feb 17, 2023

Valid Period : 6 months

Storage: 2-8° C

PRECISION

Inter-assay CV%: 6.0 7.9

Intra-assay CV%: 5.0 6.3

RECOVERY

Sample type

average% recovery

range

Serum

97


86-109%

LINEARITY

		Serum
1: 2	Average % of Expected	95
	Range (%)	90-104
1: 4	Average % of Expected	91
	Range (%)	90-106
1: 8	Average % of Expected	88
	Range (%)	87-110

STANDARD CURVE for demonstration only:

	Con ug/ml	OD
A	0	1.829
B	5	1.287
C	10	0.922
D	25	0.654
E	50	0.427
F	100	0.186



Date: 2023.02.17