

Certificate of Analysis

Product name: Human Intact Fibroblast Growth Factor 23 ELISA Kit

Catalog No.: AAA17207 **Lot No.:** 20180518

Exp: 20181118

Expiration data: 6 months

Storage: 2-8° C

PRECISION

Inter-assay CV%: 7.0 6.3

Intra-assay CV%: 5.9 6.1

RECOVERY

Sample type

average% recovery

range

Serum

97


87-107%

LINEARITY

		Serum
1: 2	Average % of Expected	96
	Range (%)	94-102
1: 4	Average % of Expected	92
	Range (%)	91-105
1: 8	Average % of Expected	88
	Range (%)	83-110

STANDARD CURVE for demonstration only:

	Con pg/ml	OD
A	0	1.962
B	50	1.537
C	100	1.185
D	250	0.655
E	500	0.314
F	1000	0.096



Date: 2018.05.18