

## Certificate of Analysis

### Mouse PGF2 $\alpha$ (Prostaglandin F2 Alpha) ELISA Kit

**Cat. No.: AAA17515**

**Size:** 96T

**Lot:** FN240308

**Principle:** This ELISA kit uses Competitive-ELISA as the method

#### Performance Characteristics

Intra-Assay CV: <8%

Inter-Assay CV: <10%

**Product expiration Date:** 2024/09/08

**Manufacture date:** 2024/03/08

Inspection Item	Standard requirements	Inspection Result
Appearance	Complete outer packing and without damage	PASS
	Correct labels	
	Correct specification instruction	
	Complete reagent, cover sealed	
	Clear liquid, no sediment or flocs	
	Sealed plate	
	Complete components	
Standard Curve	The linear correlation coefficient should not be less than 0.9900	
Coefficient of Variation	Not larger than 10%	

# for research use only

This product has been tested by Quality Control Department and meets performance specifications mentioned in the manual. (The humidity in the laboratory is 20%-60%, and the temperature is 18°C-25°C. TMB was balanced to 37°C before color development, and incubated at 37°C for 15 minutes in the dark after adding the enzyme label plate holes.)

STD.(pg/ml)	OD-1	OD-2	Average
0	2.241	2.356	2.287
7.812	1.756	1.846	1.792
15.625	1.457	1.532	1.487
31.25	1.21	1.272	1.235
62.5	0.949	0.997	0.968
125	0.748	0.786	0.763
250	0.423	0.445	0.432
500	0.318	0.334	0.324

