

Certificate of Analysis

Rat LH(Luteinizing Hormone) ELISA Kit

Cat.No.: AAA17658

Size: 96T


Lot: R1123I019

Principle: This ELISA kit uses Competitive-ELISA as the method

Product expiration Date: 2023/07/29

This product has been tested by Quality Control Department and meets performance specifications mentioned in the manual.

STD.(mIU/ml)	OD-1	OD-2	Average
0	2.071	2.131	2.101
0.312	1.665	1.713	1.689
0.625	1.38	1.42	1.4
1.25	1.033	1.063	1.048
2.5	0.749	0.771	0.76
5	0.499	0.513	0.506
10	0.241	0.247	0.244
20	0.101	0.103	0.102



Performance Characteristics

Intra-Assay CV: <8%

Inter-Assay CV: <10%

Manufacture date: 2023/01/29