

PRECISION

Intra-assay Precision (Precision within an assay): CV%<8%


Three samples of known concentration were tested twenty times on one plate to assess.

Inter-assay Precision (Precision between assays): CV%<10%

Three samples of known concentration were tested in twenty assays to assess.

TYPICAL DATA

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



pg/ml	OD1	OD2	Average
200	0.054	0.056	0.055
100	0.094	0.099	0.097
50	0.181	0.183	0.182
25	0.310	0.302	0.306
12.5	0.594	0.546	0.570
6.25	0.995	0.985	0.990
3.12	1.330	1.346	1.338
0	1.684	1.605	1.645

LOD

0.78 pg/ml

LINEARITY

To assess the linearity of the assay, samples were spiked with high concentrations of horse LEP in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.

	Sample	Serum(n=4)
1:100	Average %	99
	Range %	95-104
1:200	Average %	103
	Range %	98-109
1:400	Average %	89
	Range %	84-95
1:800	Average %	96
	Range %	91-101

RECOVERY

The recovery of horse LEP spiked to levels throughout the range of the assay in various matrices was evaluated. Samples were diluted prior to assay as directed in the Sample Preparation section.

Sample Type	Average % Recovery	Range
Serum (n=5)	93	86-100
EDTA plasma (n=4)	104	100-108