

## Certificate of Analysis

COA No:20240711025

Catalog	AAA22474
Date	2024/07/11
Lot	30361067
Quantity	1Kit

This is to certify that we, the undersigned, have inspected the quality of the above mentioned kit and found the result of inspection as follows.

Items	Result
Toxicity	None
Specific	No cross-reaction with other factors
Relative standard curve coefficient	0.9991
Intra assay Precision	$\leq 8\%$
Recovery	82%