

Certification of Analysis

Product name: Human Aflatoxin Albumin Adduct (AFT-HSA) ELISA Kit

Catalog NO.: AAA23678

Expiration date: 6 months

Storage: 2-8°C

PRECISION:

In the same lot: CV% 7.1, 6.5, 7.6

Different lot: CV% 3.6, 7.2, 4.8

RECOVERY

Sample type	average % recovery	range
Cell culture	94	88-98%
Serum	92	90-99%

LINEARITY

		Cell culture	Serum
1:2	Average % Expected	94	93
	Range (%)	92-101	89-99
1:4	Average % Expected	93	98
	Range (%)	90-104	95-103
1:8	Average % Expected	97	99
	Range (%)	90-99	91-97
1:16	Average % Expected	95	95
	Range (%)	87-98	91-101