

## Certification of Analysis

Product name: Human Aflatoxin Albumin Adduct (AFT-HSA) ELISA Kit

Catalog No.: AAA23678

Expiration date: 6 months

Storage: 2-8°C

### PRECISION:

In the same lot: CV% 3.2, 6.5, 4.9

Different lot: CV% 5.6, 7.2, 4.8

### RECOVERY

Sample type	average % recovery	range
Cell culture	94	88-97%
Serum	97	90-102%

### LINEARITY

		Cell culture	Serum
1:2	Average % Expected	94	93
	Range (%)	93-101	89-99
1:4	Average % Expected	93	98
	Range (%)	90-104	95-103
1:8	Average % Expected	97	99
	Range (%)	90-98	91-99
1:16	Average % Expected	95	95
	Range (%)	87-98	91-102