

for research use only

Certificate of Analysis

Product name: Human anti cytochrome P450c21/21 hydroxylase antibody ELISA Kit

Catalog No.: AAA27083

Lot No: 20250106C **Exp:** Jul 06,2025

Manufacture date : Jan 06,2025 **Valid Period :** 6 months **Storage:** 2-8° C

PRECISION

Inter-assay CV%: 6.3 7.8

Intra-assay CV%: 5.1 6.4

RECOVERY


Sample type	average% recovery	range
Serum	97	92-102%

LINEARITY

		Serum
1: 2	Average % of Expected	97
	Range (%)	93-100
1: 4	Average % of Expected	94
	Range (%)	87-106
1: 8	Average % of Expected	91
	Range (%)	89-112

STANDARD CURVE for demonstration only:

	Con ng/ml	OD
A	0	2.314
B	50	1.327
C	100	0.996
D	250	0.702
E	500	0.361
F	1000	0.174



Date: 2025.01.06