

## Certificate of Analysis

**Product name:** Human anti Mi2 antibody ELISAKit

**Catalog No.:** AAA27280    **Lot No:** 20240401C    **Exp:** Oct 01,2024

**Manufacture date:** Apr 01,2024    **Valid Period :** 6 months    **Storage:** 2-8° C

### PRECISION

Inter-assay CV%: 6.2 7.9    Intra-assay CV%: 5.3 6.7

### RECOVERY


| Sample type | average% recovery | range   |
|-------------|-------------------|---------|
| Serum       | 98                | 94-102% |

### LINEARITY

|      |                       | Serum  |
|------|-----------------------|--------|
| 1: 2 | Average % of Expected | 97     |
|      | Range (%)             | 94-102 |
| 1: 4 | Average % of Expected | 93     |
|      | Range (%)             | 90-104 |
| 1: 8 | Average % of Expected | 90     |
|      | Range (%)             | 86-112 |

### STANDARD CURVE for demonstration only:

|   | Con ng/mL | OD    |
|---|-----------|-------|
| A | 0         | 2.138 |
| B | 50        | 1.572 |
| C | 100       | 0.984 |
| D | 250       | 0.56  |
| E | 500       | 0.247 |
| F | 1000      | 0.161 |



Date: 2024.04.01