

for research use only

Certificate of Analysis

Product name: Rat Complement Fragment 3d, C3d ELISA

Kit Catalog No.: AAA27321

Lot No: 20231218C

Exp: Jun 18,2024

Manufacture date: Jun 18, 2023

Valid Period : 6 months

Storage: 2-8° C

PRECISION

Inter-assay CV%: 6.4 7.3

Intra-assay CV%: 5.2 6.0

RECOVERY

Sample type

average% recovery

range

Serum

96


91-101%

LINEARITY

		Serum
1: 2	Average % of Expected	97
	Range (%)	94-100
1: 4	Average % of Expected	93
	Range (%)	92-107
1: 8	Average % of Expected	92
	Range (%)	88-110

STANDARD CURVE for demonstration only:

	Con ug/mL	OD
A	0	2.056
B	0.5	1.625
C	1	1.345
D	2.5	0.906
E	5	0.515
F	10	0.238



Date: 2023.12.18