

for research use only

Certificate of Analysis

Product name: Human Semaphorin 3C ELISA Kit

CatalogNo.:AAA27332

Lot No: 20240506C

Exp:Nov 06,2024

Valid Period : 6 months **Storage:** 2-8° C

Manufacture date: May 06,2024

PRECISION

Intra-assay CV%: 5.6 6.7

Inter-assay CV%: 6.0 7.2

RECOVERY


Sample type	average% recovery	range
Serum	97	94-99%

LINEARITY

		Serum
1: 2	Average % of Expected	96
	Range (%)	93-101
1: 4	Average % of Expected	93
	Range (%)	90-107
1: 8	Average % of Expected	91
	Range (%)	88-106

STANDARD CURVE for demonstration only:

	Con ng/mL	OD
A	0	1.937
B	0.5	1.468
C	1	1.051
D	2.5	0.647
E	5	0.332
F	10	0.154



Date: 2024.05.06